



**Environmental & Technical Services Limited**  
 Unit 7 Millbrook Business Park, Hoe Lane,  
 Nazeing, Essex EN9 2RJ  
 Tel: +44 (0)1992 899 440  
 Fax: +44 (0)1992 899 441  
 Email: climatic@ets.co.uk Web: www.ets.co.uk



*Specialist Suppliers to the Environmental Industry*

## HIRE CHAMBER SPECIFICATIONS

ETS Serial No.	1072-1	Useful Capacity (Litres)	553
Chamber Type	Climatic	Manufacturer	ATT
Model No.	FM600 C	Manuf. Serial No.	TT02469
Date of Manufacture	2017	Date of Refurbishment	n/a
Internal Dimensions	850 x 730 x 892 mm (WxDxH)		
External Dimensions	1124 x 1768 x 2049 mm (WxDxH)		
Weight	1,090 kgs		
Temperature Range	-75°C to +180°C		
Temperature Rate of Change	6°C from -70°C to +180°C 4°C from +180°C to -70°C without the 'Cold Sink' device 5.5°C from +180°C to -70°C with the 'Cold Sink' device		
Humidity Range	10% to 98% RH within the temperature range +10°C to +95°C		
Altitude Range	n/a		
Power Supply Requirement	400V / 3 Phase / 32 Amps + N + E	Compressed Air Requirements	n/a
Water Requirements	Demineralised Water	Drainage Requirements	Yes
Door Hanging (L/R)	Left	Window (Y/N)	Y
		Size (mm)	W: 450 H: 550
Cable Entry Ports (No. size & location)	1 x 80mm dia. Left Hand Side 1 x 150mm dia. Right Hand Side		
Type of Controller	ATT MyKratos		
Additional Instruments	MyAngel		
Castors (Y/N)	Yes	No. of Shelves	1
Type of Cooling System	Mechanical Air Cooled Cascade	Refrigerant Used	R23 = 0.8 kg R404A = 4.5 kg
Additional Notes			



**E&OE**



VAT No. 355 1342 72  
 REGISTERED IN ENGLAND No. 1723607  
 REGISTERED OFFICE: SUITE A, 10<sup>TH</sup> FLOOR, MAPLE HOUSE, HIGH STREET, POTTERS BAR, EN6 5BS

WEEE Producer Compliance Scheme Registration No. WEE/DJ1117VX

We accept payments by

